


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


APR 30 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N		130493
Name		SMITH, PATRICK J	ID#	6397
			Date	04/22/2010
A	Agency	Clear Security Forces	Phone #	(907) 585-6293
Instrument Location		Bldg 37, Clear AFS		
Alco S/N		X124609	Target Value	.081
			High Pressure	1150
B	Alco Test Values	.075	.077	
		1 st Alco	2 nd Alco	
Signature				
(OVER)				


(Do Not write in the area below)



I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of June, 2010.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493 ✓

APRIL 22, 2010

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: CLR4

CASE/REPORT: 1234567

TEST TYPE: V

ALCO TARGET VALUE: .001 ✓

ALCO S/N: X124609 ✓

— BREATH ANALYSIS —

.001 ADJUSTED FOR 29.34 in

ALCO TARGET .079 17:00

BLANK TEST .000 17:01

INTERNAL STANDARD VERIFIED 17:01

ALCO TV 29.34 in .075 17:02

BLANK TEST .000 17:02

SUBJECT SAMPLE .000 17:03

BLANK TEST .000 17:04

ALCO TV 29.34 in .077 17:04

BLANK TEST .000 17:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493 ✓

APRIL 22, 2010 ✓

TIME 16:58

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.34 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~